

# **Aliah University**

(A UGC approved autonomous Institution under the Department of Minority Affairs and Madrasah Education, Govt. of West Bengal)

IIA/27, New Town, Rajarhat, Kolkata – 700 160, West Bengal

Web: www.aliah.ac.in

### NOTICE INVITING QUOTATION

Ref No: 009/AU/REG/NIQ/18-19 Date: 27/06/2018

Sub: Sealed Quotations are invited from the bonafide and resourceful Agents/Wholesalers/Suppliers for Supply and Installation of laboratory equipments for the Microbiology laboratory of the Dept of Biological Sciences, Aliah University, New Town Campus

Aliah University, Kolkata a Premier Educational Institute under the Minority Affairs and Madrasah Education Department, Government of West Bengal, invites <u>Sealed Quotations</u> from the bonafide and resourcefulAgents/Wholesalers/Suppliers for Supply and Installation of laboratory equipments for the Microbiology laboratory of the Dept of Biological Sciences, Aliah University, New Town Campus. The tentative quantity of the required items along with technical configuration of each items are mentioned at Annexure separately.

Interested **Bidders**may submit their **Quotation** complete in all respect To, The Registrar Aliah University IIA/27, New Town, Kolkata-700160, West Bengal, India by **13/07/2018 at 2 P.M** 

#### Scope of Work:

In this regard NIQ has been invited in two fold basis i.e. Technical Bid and Financial Bid.

Interested bidders are requested to provide their Quotes following the format in Annexure- II in their official letter heads along with signed Compliance Statement and Price Bid (Annexure- III).

The University retains the right to cancel any of the items at a later date after the contract is awarded. The University at its own discretion may cancel any or all the bids without assigning any reason thereof.

For any information in this regard please Email: <u>registrar@aliah.ac.in</u> and copy to store&purchase@aliah.ac.in

S1. Schedule	Date & Time
1 Date of uploading of NIQ (Publishing Date) at Aliah University Website	05/07/2018 at 02 P.M
3 Bid submission start date	05/07/2018 at 03 P.M
4 Bid Submission closing	13/07/2018 at 2 P.M
5 Techno Commercial Bid opening date O/o The <b>Registrar</b> , <b>Aliah</b>	16/07/2018 at 2 P.M
University, New Town (Tentative)	

Sd/-Registrar Aliah University

#### **ANNEXURE I: GENERAL TERMS & CONDITIONS**

- 1. **Supply of work**must be completed within 21 days of issuing work order and must be delivered at Dept of Biological Sciences, Aliah University, New Town Campus
- 2. The Quotationer should bear all the transportation & insurance **risk** till the on door delivery point. Selected bidder shall take all possible care for Govt. Property & of any damages due to negligence of his workers; the bidder/Agency shall be responsible for all such damages & repair the same at his own cost.
- 3. **The rates** so quoted must be inclusive of GST, Central Excise, customs Duty if any, packing freight to destination, Insurances and levies and necessary installation and fixing at designated places at Aliah University and all charges i.e. cost of Equipment and other incidental charges for supplying at destination level and onsite warranty 1 (One) year and also delivery charges up to the point of delivery at proper destination level and as per instruction in the work/ supply order. No extra charges will be

entertained. Prices can be quoted in Indian Currency only (₹).No extra payment will be made for carrying of materials involving head load/ trolley etc.

- 4. All necessaries cables and adapters for functioning of the equipments to be supply along with the Work
- 5. Supply of Items will be made in conformity with the specification & time as mentioned in the work order as decided by the authority. **No deviation in specification** will be accepted. After delivery of the materials to the respective points by selected bidder (s), authority reserves the right to collect the samples of supply the materials at random basis and send those materials for testing to ensure the quality of products etc. If it is found that materials are not according to the specification, the authority has every right to cancel the total lot or otherwise forfeit the security money, blacklisting the respective Manufacturer / Supplier and terminate the contract.
- 6. If any part of the service in respect of the work assigned and undertaken by you not rendered/delivered in time, Aliah University shall be entitled to levy and recover liquidated damages/ penalty at 1% per week or part thereof the delay/ default, subject to 5% maximum, on the payment due to the agency/contractor for the particular stage. Any delay beyond scheduled dates may attract higher penalty to be decided by the Aliah University
- 7. The bidder will be selected on overall rate only. The bidder must quote in all items otherwise their bids will be rejected.
- 8. All disputes are subject to exclusive jurisdiction of competent Court and Forum in Kolkata, India only.
- 9. Any dispute arising out of this contract shall be referred to the Registrar, Aliah University, and if either of the parties hereto is dissatisfied with the decision, the dispute shall be referred to the decision of an Arbitrator, who should be acceptable to both the parties, to be appointed by the Vice-Chancellor of the University. The decision of such Arbitrator shall be final and binding on both the parties.
- 10. Payment terms: 100% payment will be released within (30) days only after successful installation and commissioning of the supplied items duly certified by the concern Department. No advance payment or payment against Performa invoice will be made. Payment will be made after receipt, inspection, and installation/testing. The payment will be made by RTGS / FUND Transfer mode only. Advance payment not allowed. Hence, following information must be clearly written in the Price Bid for RTGS / FUND TRANSFER:
  - A. Name of the Firm with complete postal address
  - B. Name of the Bank with Branch where the Account exist
  - C. IFSC CODE
  - D. ACCOUNT No
  - E. PAN No
- 11. The Quotations are liable to be rejected if the fore going conditions are not complied with. The bid should be complete in all respects and duly signed wherever required. Incomplete and unsigned offer will not be accepted.
- 12. The products asked for should be of very high standard and of reputed brand and preferably with **B.I.S/I.S.I** code.

## Annexure II: Technical Bid Application Format (Please attach all relevant documents)

To, The Registrar Aliah University IIA/27, New Town, Kolkata-700 160

Kolkata-700 160			
		of laboratory equipments fo ah University, New Town Can	
Ref:N.I.Q. No		dated	
Sir,			
1. ABOUT THE ORGANIZATION			

	1. ABOUT THE ORGANIZATION			
1.1	Name of the Organization			
1.2	Name of Authorized Person			
1.3	Registered Office Address with telephone no. & email address			

1.4	Authorized Service Station			
	Name, address, contact person			
	name, phone number, e-mail			
	2. TECHNICAL DOCUMENTS			
2.1	Company Registration No./Trade License No./Partnership			
	Deed No. (Photocopy Required to B	e Submitted along with		
	NIQ)			
2.2	PAN Registration No			
	(Photocopy Required to Be Submitte	ed along with NIQ)		
2.3	GST Registration No			
	(Photocopy Required to Be Submitte	ed along with NIQ)		
2.4	Self Declaration that the Bidder is	Not Blacklisted by any		
	State or National Government			
	Organization, Trust, Institution, Un	iversity.		

#### ANNEXURE III COMPLIANCE STATEMENT AND PRICE BID:

Para of Quotation   Enquiry Specification of Items Offered   UNIT		ANNEXURE III COMPLIANCE STATEMENT AND PRICE BID:					
UNIT Quotation specification whether YES/ NO  1 2 3 4  Rate(Quant ity X Unit Price)  1 Laminar Air Flow:Horizontal. Working space: 4ft x 2ft x 2ft. Working table top covered with Stainless steel 304 grade sheet. Noise level below 65db.Support stand with caster wheels. Air velocity 100FPM, Outer thick ply wood covered with white summica. Inner area epoxy painted. Front door: UV sterilized polycarbonate sheet with folding type (transparent). Filter: The main filter is HEPA filter with particle retention of not more than 0.3 micron.UV lamp & Fluorescent lamp should be present.Static pressure manometer.  2 pH Meter: Digital Display, two point calibration with lockable standardization control. Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01pH pH Accuracy: ± 0.01pH mV Accuracy: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES		Para of Quotation	QTY	Compliance	TOT	AL RATI	C
Specification whether   YES/NO	S1	Enquiry Specification of Items Offered	and				
1 2 3 4    Laminar Air Flow:Horizontal. Working space: 4ft x 2ft x			UNIT	•			
1						king& In	stallation
1 Laminar Air Flow:Horizontal. Working space: 4ft ty X Unit Price)  1 Laminar Air Flow:Horizontal. Working space: 4ft x 2ft X 2ft. Working table top covered with Stainless steel 304 grade sheet. Noise level below 65db.Support stand with caster wheels. Air velocity 100FPM, Outer thick ply wood covered with white summica. Inner area epoxy painted. Front door: UV sterilized polycarbonate sheet with folding type (transparent). Filter: The main filter is HEPA filter with particle retention of not more than 0.3 micron.UV lamp & Fluorescent lamp should be present. Static pressure manometer.  2 pH Meter: Digital Display, two point calibration with lockable standardization control. Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ± 1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES					Charges		
Rate(Quant ity X Unit Price)							
1 Laminar Air Flow:Horizontal. Working space: 4ft x 2ft x 2ft. Working table top covered with Stainless steel 304 grade sheet. Noise level below 65db. Support stand with caster wheels. Air velocity 100FPM, Outer thick ply wood covered with white sunmica. Inner area epoxy painted. Front door: UV sterilized polycarbonate sheet with folding type (transparent). Filter: The main filter is HEPA filter with particle retention of not more than 0.3 micron.UV lamp & Fluorescent lamp should be present. Static pressure manometer.  2 pH Meter: Digital Display, two point calibration with lockable standardization control. Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01pH pH Accuracy: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES		1	2	3			1
1 Laminar Air Flow:Horizontal. Working space: 4ft x 2ft x 2ft. Working table top covered with Stainless steel 304 grade sheet. Noise level below 65db.Support stand with caster wheels. Air velocity 100FPM, Outer thick ply wood covered with white sunmica. Inner area epoxy painted. Front door: UV sterilized polycarbonate sheet with folding type (transparent). Filter: The main filter is HEPA filter with particle retention of not more than 0.3 micron.UV lamp & Fluorescent lamp should be present.Static pressure manometer.  2 PH Meter: Digital Display, two point calibration with lockable standardization control. Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01 pH pH Resolution: ± 0.01 pH pH Accuracy: ±0.01 pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES						GST	Total
Laminar Air Flow:Horizontal. Working space: 4ft x 2ft x 2ft. Working table top covered with Stainless steel 304 grade sheet. Noise level below 65db.Support stand with caster wheels. Air velocity 100FPM, Outer thick ply wood covered with white sunmica. Inner area epoxy painted. Front door: UV sterilized polycarbonate sheet with folding type (transparent). Filter: The main filter is HEPA filter with particle retention of not more than 0.3 micron.UV lamp & Fluorescent lamp should be present. Static pressure manometer.  2 pH Meter: Digital Display, two point calibration with lockable standardization control. Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ± 1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES					•		
x 2ft x 2ft. Working table top covered with Stainless steel 304 grade sheet. Noise level below 65db. Support stand with caster wheels. Air velocity 100FPM, Outer thick ply wood covered with white summica. Inner area epoxy painted. Front door: UV sterilized polycarbonate sheet with folding type (transparent). Filter: The main filter is HEPA filter with particle retention of not more than 0.3 micron.UV lamp & Fluorescent lamp should be present. Static pressure manometer.  2 pH Meter: Digital Display, two point calibration with lockable standardization control. Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01pH pH Accuracy: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES	1	Tomboo Ata Disamiliani and all Windian and a	46 1		Unit Price)		
Stainless steel 304 grade sheet. Noise level below 65db. Support stand with caster wheels. Air velocity 100FPM, Outer thick ply wood covered with white sunmica. Inner area epoxy painted. Front door: UV sterilized polycarbonate sheet with folding type (transparent). Filter: The main filter is HEPA filter with particle retention of not more than 0.3 micron.UV lamp & Fluorescent lamp should be present.Static pressure manometer.  2 pH Meter: Digital Display, two point calibration with lockable standardization control. Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01pH pH Accuracy: ±0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES	1						
65db.Support stand with caster wheels. Air velocity 100FPM, Outer thick ply wood covered with white sunmica. Inner area epoxy painted. Front door: UV sterilized polycarbonate sheet with folding type (transparent). Filter: The main filter is HEPA filter with particle retention of not more than 0.3 micron.UV lamp & Fluorescent lamp should be present.Static pressure manometer.  2 pH Meter: Digital Display, two point calibration with lockable standardization control. Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
velocity 100FPM, Outer thick ply wood covered with white sunmica. Inner area epoxy painted. Front door: UV sterilized polycarbonate sheet with folding type (transparent). Filter: The main filter is HEPA filter with particle retention of not more than 0.3 micron.UV lamp & Fluorescent lamp should be present.Static pressure manometer.  2 pH Meter: Digital Display, two point calibration with lockable standardization control. Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01pH pH Accuracy: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
with white sunmica. Inner area epoxy painted. Front door: UV sterilized polycarbonate sheet with folding type (transparent). Filter: The main filter is HEPA filter with particle retention of not more than 0.3 micron.UV lamp & Fluorescent lamp should be present.Static pressure manometer.  2 pH Meter: Digital Display, two point calibration with lockable standardization control. Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01 pH pH Accuracy: ± 0.01 pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
Front door: UV sterilized polycarbonate sheet with folding type (transparent). Filter: The main filter is HEPA filter with particle retention of not more than 0.3 micron.UV lamp & Fluorescent lamp should be present.Static pressure manometer.  2 pH Meter: Digital Display, two point calibration with lockable standardization control.  Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01pH pH Accuracy: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
with folding type (transparent). Filter: The main filter is HEPA filter with particle retention of not more than 0.3 micron.UV lamp & Fluorescent lamp should be present.Static pressure manometer.  2 pH Meter: Digital Display, two point calibration with lockable standardization control. Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01pH pH Accuracy: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
filter is HEPA filter with particle retention of not more than 0.3 micron.UV lamp & Fluorescent lamp should be present.Static pressure manometer.  2 pH Meter: Digital Display, two point calibration with lockable standardization control. Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01pH pH Accuracy: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
lamp should be present. Static pressure manometer.  2 pH Meter: Digital Display, two point calibration with lockable standardization control. Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01 pH pH Accuracy: ± 0.01 pH mV Range: 0 to + 1999 MV mV Accuracy: ±1 mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
manometer.  2 pH Meter: Digital Display, two point calibration with lockable standardization control. Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01pH pH Accuracy: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES		more than 0.3 micron.UV lamp & Fluorescent					
pH Meter: Digital Display, two point calibration with lockable standardization control. Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01pH pH Accuracy: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
with lockable standardization control.  Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01pH pH Accuracy: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
Display: 4 digit LED pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01pH pH Accuracy: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES	2	J 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1					
pH Range: 0.00 to 14.00 pH pH Resolution: ± 0.01pH pH Accuracy: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
pH Resolution: ± 0.01pH pH Accuracy: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
pH Accuracy: ± 0.01pH mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
mV Range: 0 to + 1999Mv mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
mV Accuracy: ±1mV Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
Temp: Manual Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
Electrode: Unbreakable  3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
3 Magnetic Stirrer: 2 liter capacity with hot plate and Digital Speed Indicator  TOTAL QUOTE IN RUPEES							
and Digital Speed Indicator  TOTAL QUOTE IN RUPEES	3		ite 1pc				
TOTAL QUOTE IN WORD		TOTAL QUOTE IN RUPEES					
		TOTAL QUOTE IN WORD					

conditions as mentioned in the invitation letter. We hereby certify that we have taken steps to ensure that no person acting for us or on our behalf will engage in bribery. Certify that all above information are correct to the best of my/our information, knowledge and belief.

Signature of the Bidder Name	Date
Designation Seal	
Ref. No: 009/AU/REG/NIQ./18-19	Dated: 27/06/2018

Copy to:

- 1. Chairman, Departmental Purchase Committee
- 2. HoD, Deptt of Biological Sciences
- 3. Notice Board at Aliah University
- 4. Website: www.aliah.ac.in
- 5. Guard File

Sd/-Registrar